Check list for wall suction devices in Hywel Dda University Health Board

Low suction use area (e.g. general wards)

Ceredig Beds 16-19

Check should be undertaken daily or between patients (whichever is the highest frequency)

Check that

- ✓ Device, tubing and canister are clean and free from dust.
- ✓ Filter not discoloured (change if needed) and in date.
- ✓ All tubing attached correctly, unused yankuer sucker attached but kept in plastic sleeve.
- ✓ External canister in good condition and in date.
- ✓ Liner empty and in date.
- ✓ Function properly (suction needle moves to at least halfway up the dial)

Filter expiry date should be 1 year from date it was changed.

Sign below to confirm that above checks were completed and remedial action taken where necessary for each device.

Device Date	Bed 16	Bed 17	Bed 18	Bed 19	Bed 20	Bed 21	Bed 22	Bed 23	Bed 24	Bed 25	Bed 26	Bed 27	Bed 28	Bed 29	Bed 30
Date	Problems							Remedial							