

Micromed Electronics Ltd

11 Drakes Way,
Mayford,
Woking,
Surrey,
GU22 0NX
Tel : 01483 728822

www.micromeduk.com

Planned Service Visit Report Form (MPS01)

Hospital / Dept : Hafan Derwen, ECT Department.

Equipment Model / Mecta Spectrum 5000M / Ser No 10728

Year of Manufacture : 2002

CE029098

Item No Equipment Test

1	Check mains plug wiring integrity.	✓
2	Check mains plug for correct fuses.	✓
3	Check stimulus paddles condition / function.	✓
4	Check monitoring leads condition / function.	✓
5	Check touch panel condition / function.	✓
6	Record central processor firmware version.	VI-2-3 GEMSO
7	Record safety processor firmware version.	VI-1-9
8	Check all stimulus parameters function.	✓
9	Check printer condition / function.	✓
10	Check stimulus paddles impedance function.	✓
11	Check stimulus output function.	✓
12	Electrical Safety Test	✓

Engineer... Coli Peltat ... Date... 7/8/19 ...

Customers Signature... E. Daines ...