

CE010157

Micromed Electronics Ltd
11 Drakes Way
Mayford
Woking
Surrey
GU22 0NX
Tel: 01483 728822

www.micromeduk.com

Planned Service Visit Report Form (MPS01)

Hospital / Dept : Hafan Derwen, ECT Department.

Equipment Model / Mecta Spectrum 5000M / Ser No 13083
13083

Year of Manufacture : 2017
CE010157

Item No Equipment Test

1	Check mains plug wiring integrity.	✓
2	Check mains plug for correct fuses.	✓
3	Check stimulus paddies condition / function.	✓
4	Check monitoring leads condition / function.	✓
5	Check touch panel condition / function.	✓
6	Record central processor firmware version.	V1.4.0BMS0A
7	Record safety processor firmware version.	V1.2.7
8	Check all stimulus parameters function.	✓
9	Check printer condition / function.	✓
10	Check stimulus paddies impedance function.	✓
11	Check stimulus output function.	✓
12	Electrical Safety Test	✓

Engineer... *Colin Pettit* ... Date... 7/8/19

Customers Signature... *E. Daines*